

# Expert information

# **FOKALDRY**

Manufacturer (€ 0124

lege artis Pharma GmbH + Co. KG P. O. Box 60, D-72132 Dettenhausen Breitwasenring 1, D-72135 Dettenhausen Telephone +49 (0) 71 57 / 56 45 - 0 Fax +49 (0) 71 57 / 56 45 50 E-Mail: info@legeartis.de www.legeartis.de

#### Name

**FOKALDRY** 

Cleaning solution for the cavity, the root canal, and the tooth surface. For dental use only

## Composition

Acetone, isopropyl alcohol, petroleum ether, and diisopropylether

#### **Indications**

Cleaning, degreasing and drying of cavities, tooth stumps, root canals, crowns bridges and brackets. Removal of plaque on the hard substance of the tooth.

Removal of uncured cement and composite remnants.

### Side effects

In case of gingival or mucous membrane contact, burning may occur, avoid contact as much as possible. Inhaling FOKALDRY may cause irritation of the throat and pharyngx. When inhaling larger quantities, headache, nausea, and sensation of dizziness may occur.

In rare cases, repeated contact may lead to an allergic reaction to isopropyl alcohol.

### **Notices and precautions**

Avoid swallowing. After swallowing, rinse mouth. Do not induce vomiting. In case of vomiting danger of aspiration. Seek medical advice/medical attention immediately.

In case of skin contact, wash with plenty of water and soap. If eye contact occurs, rinse with water for a few minutes, remove any existing contact lenses if possible.

The safety data sheet can be downloaded from <u>www.legeartis.de</u> or requested by E-Mail from <u>sicher-heitsdatenblaetter@legeartis.de</u>.

# Mode of application

A cotton wool pellet of appropriate size is soaked with FOKALDRY and used to clean the cavity resp. the stump.

Crowns, bridges and brackets should be cleaned with FOKALDRY prior their permanent fixation. In case of plastic crowns, partial crowns or brackets embrittlements or matting during application may occur depending on the used plastic. In this regard, the information from the respectively manufacturer regarding material compatibility has also to be considered. Moisten the root canal with FOKALDRY, and dry it with paper points. After evaporation FOKALDRY does not leave a smear film.

Take care that the liquid does not come over the apex into the periapical tissue, while cleaning the root canal.

As far as possible a rubber dam is to be used.

# **Further notices**

FOKALDRY serves to remove fat-containing remainders of tooth preparations and humidity to achieve a better adaption resp. compound with the rebasing resp. filling material and anchorage cements. Too extensive use of FOKALDRY may cause dehydration of dentin and upset the physical balance of interstitial liquid of the pulp.

Do not pour back residues of the solution into the bottle.

## Shelf life

FOKALDRY shall not be used after the expiry date.

## Storage and preservation

Close the bottle immediately after use and keep protected from light and heat. FOKALDRY is an easily inflammable liquid, therefore, do not handle or store near sources of ignition. In the case of larger quantities of evaporated liquid, ensure sufficient aeration.

# Administrative form and package sizes

80 ml solution ltem number 0032305

500 ml solution (480-ml reserve bottle with additional spout for decanting and 20 ml application bottle with 3 drop dispensers for easy dropwise removal) Item number 0032306

#### Date of revision

2019-11

Symbol of "Expiry date"

Symbol of "Batch number"

Symbol of "Pay attention to the directions for use"



LOT



Symbol of "Manufacturer"

Symbol of "Catalogue number"
Shows Item number of the manufacturer



